

FBG Water-Level Transducer Based on PVC-Cantilever and Rubber-Diaphragm Structure

E. Vorathin¹, Z. M. Hafizi¹, N. Ismail¹, M. Loman¹, A. M. Aizzuddin¹, K. S. Lim² and H. Ahmad²

¹Advanced Structural Integrity and Vibration Research, Faculty of Mechanical & Manufacturing Engineering, Universiti Malaysia Pahang, Pekan 26600, Malaysia

(e-mail: vora.91.11@hotmail.com; hafizi@ump.edu.my; nurazima@ump.edu.my; miminorazeen@ump.edu.my; muhammad.aizzuddin.a@gmail.com)

²Photonics Research Centre, Faculty of Science, University of Malaya, Kuala Lumpur 50603, Malaysia

(e-mail: kslim@um.edu.my; harith@um.edu.my)

ABSTRACT

The use of FBG sensors in water level monitoring has shown great advantages over the use of conventional electrical sensors. However, much effort is still required to increase their sensitivity towards water level measurement with eliminated sensitivity towards temperature. Here, the FBG was bonded on a PVC polymer plastic cantilever beam and rubber diaphragm structure. The sensing mechanism was based on the differential FBG method. Variations of the water level from 0 to 100 cm obtained the sensitivity of 0.0639 nm/cm with the linearity of 98.06%. The poor linearity was mainly due to the manual fabrication errors. However, this uncertainty can be overcome through the polynomial regression fit in determining the water level measurement. The measured water level was closed to the actual water level at a tolerance of ± 2 cm. For temperature effects, the sensitivity was inconspicuous, recorded at only 0.0003 nm/ $^{\circ}$ C across the range of 10 $^{\circ}$ C to 60 $^{\circ}$ C.

KEYWORDS: Fiber Bragg grating (FBG), water level, temperature compensation.

DOI: 10.1109/JSEN.2019.2916469

ACKNOWLEDGMENT

The authors would like to thank the Faculty of Mechanical and Manufacturing Engineering, Universiti Malaysia Pahang (<http://www.ump.edu.my/>) for providing the laboratory facilities; and the Institute of Postgraduate Studies (IPS), Universiti Malaysia Pahang for providing the financial support through Doctoral Research Scheme (DRS) scholarship. They would also like to thank the Photonics Research Centre University Malaya (PRCUM) for the support in fabricating the FBGs.